



Certificate

ICR Polska/CE/V/ [REDACTED]

Name and address of certificate owner: [REDACTED]

Name and address of manufacturer:

Same as above

Product name:

Infrared Thermometer

Product types:

[REDACTED] T008

Product trademark: [REDACTED]

- This certificate confirms that the product meets the requirements of the following standard
- The conformity was demonstrated based on following standard(s) listed by European Commission as harmonized with **Directive 93/42/EEC**

EN 60601-1:2006+A1:2013
+A12:2014

EN 60601-1-2: 2015

The certification has been carried out in accordance with individual rules and conditions agreed with the applicant. Evaluation has been carried out in accordance with:

Test report(s) No.:

18250IC00001701E, 18250IC00001701S

Test conducted by: [REDACTED]

Certificate issue date:

03.04.2020

Expiration date:

02.04.2025

NOTES:

- This certificate refers to the above mentioned product and its conformity in regards of above mentioned standard(s) was proven on test sample
- This certificate does not imply meeting all essential requirements, assessment of the series-production or any other restricted Notified Bodies conformity assessment procedure appropriate for the product
- This certificate holder shall use this certificate in connection to declaration of conformity and technical data relevant for the product the certificate was issued

Certificate modifications: none

Director: Rafal Kalinowski

Warsaw, 03.04.2020

